


One Min Manual start



**Task**

Task ID: 13  
 Task Description: One Min Manual start  
 Task status: Active  
 Model: N2003/2013 Issue 2.00  
 Serial number: 08070061  
 Channels: 1,2  
 Sessions: 8  
 Logging mode: Manual start/stop  
 Logging interval: 1 min  
 Created by: ant 10 Jun 2009 1:43:23 pm  
 Modified by: ant 10 Jun 2009 1:44:01 pm  
 Miscellaneous: Externally created [No], Enable alarms [No], Auto-reset alarms [No], Enable bleeper [No], Memory wraparound [No], Set clock [No]

Ch	Channel name	Sensor type	Probe type	Alarm low	Alarm high	Alarm Delay	Min	Graph min	Max	Graph max	W/color
1	Humidity 1	Humidity	Standard	0.0 rh	0.0 rh	0 secs	0.0 rh	Auto (rh)	100.0 rh	Auto (rh)	1/0000FF
2	Generic Temp 2	Generic Temp	Standard	0.0 °C	0.0 °C	0 secs	-80.0 °C	Auto (°C)	150.0 °C	Auto (°C)	1/FF0000

**Session**

Session ID: 8  
 Session Description:  
 Data start/end: 6/24/2009 3:37:01 PM - 6/29/2009 1:50:01 PM  
 No. of records: 7094

**Statistics**

Channel	Minimum	Maximum	Average	Std Dev.	Below Alarm Low	Above Alarm High
Ch 1 = Humidity 1 (rh)	36.3	66.0	59.31	4.2862	0s	4 days 22 hrs 13 ...
Ch 2 = Generic Temp 2 (°C)	18.3	23.9	20.90	1.3766	0s	4 days 22 hrs 13 ...